

Certificate of Test

Equipment Under Test

Manufacturer: Processautomatic Ratex AB
Type of product: Flowmeter
Model number: EZ-Mag 20/2 & EZ-Mag 20/2 D
Serial Number: BK Test object number 10042, 10043
Tested on behalf of: Processautomatic Ratex AB

Summary

The equipment under test described in this report

complies with does not comply with

- Emission according to EN 61326-1:2013, Edition 3
Electrical equipment for measurement, control and laboratory use.
- Immunity according to EN 61326-1:2013, Edition 3 Industrial Requirements
Electrical equipment for measurement, control and laboratory use.

The test results in this report only apply to the specified sample of EUT.

Exclusions

No exclusions were made.

Date: December 10th, 2015

Issued by:


Johan Bergstrand

Approved by:


Michael Sohlberg